

Reprocessed by



Sustainability Solutions

Instructions for Use

Reprocessed DeRoyal® Chair Sensor Pad

Reprocessed Device for Single Patient Use

Caution: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

- NON-STERILE
- Exposed to Ethylene Oxide (EO) gas
- Not made with natural rubber latex

Explanation of Symbols



Date of Reprocessing



Do Not Reuse



See Instructions for Use

Reprocessed DeRoyal® Chair Sensor Pad Description

The Reprocessed DeRoyal® Chair Sensor Pads are sensor pads designed for use with the DeRoyal® Fall Monitors to alert staff and patients in the event the patient attempts to exit the bed, wheelchair and/or recliner.

Indications for Use

The Stryker Sustainability Solutions Reprocessed DeRoyal® Chair Sensor Pads alert staff and patients in the event the patient attempts to exit the bed, wheelchair and/or recliner. It is not claimed that the device will stop elopement and/or stop falls. The device is designed to augment caregiver's comprehensive resident mobility management program.

Warnings

- FAILURE TO COMPLY WITH WARNING MAY RESULT IN INJURY. The device is not suitable for all individuals. Other devices may be required. The device is not a substitute for visual monitoring for caregiver. It is not claimed that this device will stop elopement and/or stop falls. Sensor pad and fall monitor systems are prone to difficulties. Be sure that the cord cannot entangle or choke the user. The device is designed to augment caregivers' comprehensive resident mobility management program. Test the device before each use. Read the instructions including warnings before each use.
- The Stryker Sustainability Solutions Reprocessed DeRoyal® Chair Sensor Pad for use with the DeRoyal® Fall Monitor System is designed to be installed by the end-user. As such, it is the entire responsibility of the buyer to ensure that the system is properly installed and tested. Further, the system is not designed to replace good caregiving practices including but not limited to:
 - a. Direct patient supervision
 - b. Adequate training for staff personnel for resident fall prevention and elopement
 - c. Testing of the system before each use

Storage and Handling

- Protect the device from light sources; keep away from sunlight.
- Protect the device from moisture; keep dry.

Set-Up Instructions

1. Install new batteries into the DeRoyal fall monitor. The battery compartment door is secured with the small Phillips-head screw.
2. Plug the cord end of the sensor pad into the bottom of the DeRoyal fall monitor. Pinch the cord together and loop it through the strain relief recess on the fall monitor (near plug input).
3. Test the system. Put pressure on the sensor pad to activate the pad and fall monitor. The fall monitor will beep two times to indicate that the sensor pad and fall monitor are in use. The "In Use" or "Status" light will blink when pressure is on the pad. When pressure is removed, the fall monitor will sound the alert tone. Press the reset button to silence the alert.
4. Mount the fall monitor on the chair. Position the sensor pad on the chair seat where the patient will sit. While the patient sits on the chair, the system is in use. When the patient gets up from the chair, the alert will sound.

Maintenance

- When using a sensor pad in conjunction with foam or gel cushions, test the systems to make sure it functions properly
- To minimize spreading infection, it is recommended to use a new sensor pad for each patient.
- To clean sensor pads, use disinfectant wipes or anti-bacterial cleaners.
- DO NOT fold the sensor pad.
- DO NOT immerse the sensor pad in any liquid cleaning solution.
- DO NOT expose the sensor pad to high temperature sanitizers.

Troubleshooting

1. Make sure that the sensor pad is plugged into the fall monitor.
2. Check the power supply. If you are using battery power, install new batteries. Or, if you are using an AC Adapter, make sure that the adapter is plugged into a power outlet.
3. Check monitor switches:
 - a. When using a monitor with a Safety Auto Reset™ button. For normal operation (in Reset mode) the switch on the left-hand side of the fall monitor must be in "Reset Button On" position. To silence the fall monitor, press the "Reset" button or place pressure on the sensor pad (to automatically reset the monitor).
 - b. If you are using a fall monitor with an On/Off switch (on left side of the monitor), make sure the switch is in the "On" position.
4. If the steps above do not resolve your problem, call manufacturer for technical support. Make sure both the sensor pad and the fall monitor are with you at the time of the call.

Warranty

MP2200-CP - This product is warranted for up to 30 days from recorded "in-use date," in accordance with Stryker's standard limited warranty.

Reprocessed Products

Unless agreed otherwise in writing, Stryker warrants all reprocessed products, subject to the exceptions provided herein, to be free from defects in reprocessing and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for one use in accordance with the instructions for use of such product.

STRYKER SHALL NOT BE LIABLE FOR ANY DAMAGES TO THE EXTENT CAUSED BY ANY DEFECT IN MATERIAL, WORKMANSHIP OR DESIGN BY THE ORIGINAL MANUFACTURER OF THE PRODUCT OR ANY ACT OR OMISSION OF THE ORIGINAL MANUFACTURER OF THE PRODUCT.

Products for which Stryker is the Original Manufacturer

Stryker warrants all products for which it is the original manufacturer, subject to the exceptions provided herein, to be free from defects in design, materials and workmanship and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for a period of one year from the date of purchase.

General Warranty Terms Applicable to All Products

TO THE FULLEST EXTENT PERMITTED BY LAW, THE EXPRESS WARRANTY SET FORTH HEREIN IS THE ONLY WARRANTY APPLICABLE TO THE PRODUCTS AND IS EXPRESSLY IN LIEU OF ANY OTHER WARRANTY BY STRYKER, EXPRESSED OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OR MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. IN NO EVENT WILL STRYKER'S LIABILITY ARISING IN CONNECTION WITH THE SALE OF THE PRODUCT (WHETHER UNDER THE THEORIES OF BREACH OF CONTRACT, TORT, MISREPRESENTATION, FRAUD, WARRANTY, NEGLIGENCE, STRICT LIABILITY OR ANY OTHER THEORY OF LAW) EXCEED THE PURCHASE PRICE, CURRENT MARKET VALUE OR RESIDUAL VALUE OF THE PRODUCTS, WHICHEVER IS LESS. STRYKER SHALL NOT BE LIABLE FOR INDIRECT, SPECIAL, INCIDENTAL, PUNITIVE OR CONSEQUENTIAL DAMAGES RESULTING FROM ANY BREACH OF WARRANTY OR UNDER ANY OTHER LEGAL THEORY.

This warranty shall apply only to the original end-user purchaser of products directly from Stryker or a Stryker authorized distributor. This warranty may not be transferred or assigned without the express written consent of Stryker.

This warranty does not apply to: (1) products that have been misused, neglected, modified, altered, adjusted, tampered with, improperly installed or refurbished; (2) products that have been repaired by any person other than Stryker personnel without the prior written consent of Stryker; (3) products that have been subjected to unusual stress or have not been maintained in accordance with the instructions in the user manual or as demonstrated by a Stryker representative; (4) products on which any original serial numbers or other identification marks have been removed or destroyed; (5) products that have been repaired with any unauthorized or non-Stryker components; or (6) failure to follow the instructions for use.

If a valid warranty claim is received within thirty (30) days of the expiration of the applicable warranty period, Stryker will, in its sole discretion: (1) replace the product at no charge with a product that is at least functionally equivalent to the original product or (2) refund the purchase price of the product. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property. In any event, Stryker's liability for breach of warranty shall be limited to the replacement value of the defective or non-conforming part or component.

If Stryker determines in its reasonable discretion that the claimed defect or non-conformance in the product is excluded from warranty coverage as described hereunder, it will notify the customer of such determination and will provide an estimate of the cost of repair of the product. In such an event, any repair would be performed at Stryker's standard rates.

Products and product components repaired or replaced under this warranty continue to be warranted as described herein during the initial applicable warranty period or, if the initial warranty period has expired by the time the product is repaired or replaced, for thirty (30) days after delivery of the repaired or replaced product. When a product or component is replaced, the item provided in replacement will be the customer's property and the replaced item will be Stryker's property. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property.

The OEM information listed on the label is provided as device ID prior to reprocessing and may contain the trademarks of unrelated third parties that do not sponsor this device.

This product and its packaging have been exposed to ethylene oxide gas (EtO). Even though the product then is processed in compliance with all applicable laws and regulations relating to EtO exposure, Proposition 65, a State of California voter initiative, requires the following notice:

Warning: This product and its packaging have been exposed to ethylene oxide. The packaging may expose you to ethylene oxide, a chemical known to the State of California to cause cancer or birth defects or other reproductive harm.

Warning: This product can expose you to chemicals including bisphenol A, which is known to the State of California to cause birth defects or other reproductive harm. For more information, go to www.P65Warnings.ca.gov.

DeRoyal® is a registered trademark of DeRoyal Industries.

HAS EL10117 Rev. A 06/2020