

Reprocessed by

stryker

Instructions for use

**Reprocessed LigaSure Blunt Tip Laparoscopic Sealer/Divider
Without Nano-coating (Model LF1837) – Manufactured
DECEMBER 15, 2023 or Later**

Reprocessed device for single use

Caution: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

- STERILE
- NOT MADE WITH NATURAL RUBBER LATEX






LF1837

Compatible with:

FORCETRIAD SW v3.6 – v4.0

VLFT10GEN SW v1.1 – v2.1.0.14, v4.0.1.15, v4.0.2.25, v4.0.3.20 and v4.0.4.5

Explanation of symbols

Symbol	Rules/ Standard Reference	ISO 7000 Registration Number	Symbol Title	Description
	ISO 15223-1 Clause 5.1.3	2497	Date of manufacture	Indicates the date when the medical device was manufactured.
	ISO 15223-1 Clause 5.2.3	2501	Sterilized using ethylene oxide	Indicates a medical device that has been sterilized using ethylene oxide.
	ISO 15223-1 Clause 5.1.4	2607	Use-by date	Indicates the date after which the medical device is not to be used.
	ISO 15223-1 Clause 5.4.3	1641	Consult instructions for use	Indicates the need for the user to consult the instructions for use.
	ISO 15223-1 Clause 5.4.2	1051	Do not re-use	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.

Reprocessed LigaSure Blunt Tip Laparoscopic Sealer/Divider without Nano-coating Description

The Reprocessed LF1837 is designed for use with Covidien™ electrosurgical generators that include vessel sealing capability. Please refer to the cover page for details on compatible generator models and software versions. If the software version on your generator is lower than required, contact Covidien about software upgrades.

These instructions assume that the operator is knowledgeable about correct set-up and operation of the associated Covidien generator. Refer to the generator user's guide for set up information and for additional warnings and cautions.

The instrument creates a seal by application of radiofrequency (RF) electrosurgical energy to vascular structures (vessels and lymphatics) or tissue bundles interposed between the jaws of the instrument. A blade within the instrument is surgeon actuated to divide tissue.

Maximum rated voltage: 288 V_{peak}



Defibrillation-proof type CF applied part

Indications for Use


The Reprocessed LigaSure Sealer/Divider is a bipolar electrosurgical instrument intended for use in minimally invasive or open surgical procedures where ligation and division of vessels, tissue bundles, and lymphatics is desired. The LigaSure Sealer/Divider can be used on vessels (arteries and veins) up to and including 7 mm. It is indicated for use in general surgery and such surgical specialities as urologic, vascular, thoracic, and gynecologic. Procedures may include, but are not limited to, Nissen fundoplication, colectomy, cholecystectomy, adhesiolysis, hysterectomy, oophorectomy, etc.

The LigaSure system has not been shown to be effective for tubal sterilization or tubal coagulation for sterilization procedures. Do not use the LigaSure system for these procedures.

General Warnings

- The nano-coating feature on the original device is not present on this product.
- This product cannot be adequately cleaned and/or sterilized by the user in order to facilitate safe reuse, and is therefore intended for single use. Attempts to clean or sterilize these devices without appropriate regulatory authorization may result in bio-incompatibility, infection, or product failure risks to the patient.
- The instrument is intended for use ONLY with the Covidien equipment listed on the cover of this document. Use of this instrument with other generators may not result in the desired tissue effect, may result in injury to the patient or surgical team, or may cause damage to the instrument.
- Do not use the LigaSure system unless properly trained to use it in the specific procedure being undertaken. Use of this equipment without such training may result in serious unintended patient injury.
- The contents of the package are provided sterile by method of ethylene oxide gas and is for single patient use only. Do not use if there is any evidence of damage to the package.
- Use the system with caution in the presence of internal or external pacemakers or other implanted devices. Interference produced by electrosurgical equipment can cause a pacemaker or other device to enter an unsafe mode or permanently damage the device. Consult the device manufacturer or responsible hospital department for further information when use is planned in patients with implanted medical devices.
- When this instrument is used with an energized endoscope, the leakage current from the instrument and the endoscope are additive. The patient may be exposed to unexpected levels of leakage current if this instrument is used with an energized endoscope that is not type CF.
- In minimally invasive surgery, inspect the outer surfaces of the instrument before insertion through the cannula to ensure that there are no rough or sharp edges that could damage tissue.
- Contact between an active instrument electrode and any metal objects (hemostats, staples, clips, retractors, etc.) may increase current flow and may result in unintended surgical effects, such as an effect at an unintended site or insufficient energy deposition.

- The safe and effective use of RF energy depends on many factors solely under the control of the operator. There is no substitute for properly trained and vigilant personnel. It is important that the operating instructions supplied with this or any other medical equipment be read, understood, and followed.



Directions for Use

Set Up

Warning

- **Electrical Shock Hazard** – Do not connect wet accessories to the LigaSure system.
- Position instrument cords to avoid contact with the patient or other cords. Do not wrap cords around metal objects. This may induce currents that could lead to shocks, fires, or injury to the patient or surgical team.
- Examine all LigaSure system and instrument connections before using. Improper connection may result in arcing, sparks, accessory malfunction, or unintended surgical effects.
- Inspect the instrument and cords for breaks, cracks, nicks, or other damage before use. Failure to observe this caution may result in injury or electrical shock to the patient or surgical team or cause damage to the instrument. If damaged, do not use.
- Confirm proper LigaSure system settings before proceeding with surgery.
- Do not use in the presence of flammable anesthetics or oxidizing gases, such as nitrous oxide (N₂O) and oxygen, or in close proximity to volatile solvents (such as ether or alcohol) as explosion may occur.
- Because of concerns about the carcinogenic and infectious potential of electrosurgical by-products (such as tissue smoke plume and aerosols), protective eye wear, filtration masks, and effective smoke evacuation equipment should be used in both open and minimally invasive procedures.
- Connect adaptors and accessories to the electrosurgical unit only when the unit is off or in standby mode. Failure to do so may result in injury or electrical shock to the patient or operating personnel.
- In case the instrument is accidentally dropped on the floor, turn off the generator and unplug the instrument prior to discarding.

Precaution

- Inspect packaging before opening. The contents of the package are sterile if the package has not been compromised. Do not use the instrument if the sterility has been compromised. If the package is damaged or if it was opened and the instrument not used, return the instrument and the package to Stryker.
- Do not attempt to resterilize. Stryker will not accept instruments for reprocessing that have been reprocessed and sterilized by other facilities.
- If the generator provides multiple power settings, use the lowest power needed to achieve the intended effect.

1. Remove the instrument from tray by firmly pulling on the handle (4). Do not pull on the instrument's jaws (1) or cable (7). Place the instrument in a sterile work area using aseptic technique.
2. Insert the connector (6) into the receptacle on the generator. Follow the instructions in the generator user's guide to complete the setup procedure.

During Surgery

Tissue Manipulation and Dissection

The instrument can be used to manipulate and dissect tissue with the jaws either open or closed.

Warning

- Avoid placing fingers between the lever and the handle, between the lever and the trigger, or in the jaws. Injury to the user may result.
- Use caution when handling the instrument between uses to avoid accidental activation of the LigaSure system. Do not place the instrument on the patient or drapes when not in use.
- Keep the cord free from the jaw and latch area of the instrument.
- **Fire Hazard** – Do not place instruments near or in contact with flammable materials (such as gauze, surgical drapes, or flammable gases). Instruments that are activated or hot from use may cause a fire. When not using instruments, place them in a clean, dry, highly visible area not in contact with the patient. Inadvertent contact with the patient may result in burns.
- For minimally invasive procedures, be alert to these potential hazards:
 1. Do not use hybrid trocars that are comprised of both metal and plastic components. Capacitive coupling of RF current may cause unintended burns.
 2. Use the appropriately sized trocar to allow for easy insertion and extraction of the instrument.
 3. Carefully insert and withdraw the instrument through the cannula to avoid damage to the device and/or injury to the patient.
 4. Close jaws using device lever before insertion/extraction in the trocar.

Rotating the Instrument Jaws

Notice – Do not turn the rotation wheel (3) when the lever (8) is latched. Product damage may occur.

Turn the rotation wheel on the instrument until the jaws are in the required position.

Sealing Vessels and Tissue Bundles

1. Open the jaws by pushing forward on the lever.
2. Grasp the intended vessel and/or tissue in the center of the jaws.

Warning

- Do not use this instrument on vessels larger than 7 mm in diameter.
- If the instrument shaft is visibly bent, discard and replace the instrument. A bent shaft may prevent the instrument from sealing or cutting properly.
- Do not place the vessel and/or tissue in the jaw hinge. Place the vessel and/or tissue in the center of the jaws.
- Conductive fluids (e.g., blood or saline) in direct contact with, or in close proximity to, the instrument may carry electrical current or heat, which may cause unintended burns to the patient. Aspirate fluid from around the instrument jaws before activating the instrument.
- Keep the external surface of the instrument jaws away from adjacent tissue while activating the LigaSure system or unintended injury may result.
- During a seal cycle, energy is applied to the tissue between the instrument jaws. This energy may cause water to be converted into steam. The thermal energy of steam may cause unintended injury to tissues in close proximity to the jaws. Care should be taken in surgical procedures occurring in confined spaces in anticipation of this possibility.
- Eliminate tension on the tissue when sealing and cutting to ensure proper function.

- Use caution when grasping, manipulating, sealing, and dividing large tissue bundles.
 - Do not bend instrument shaft (2).
 - Do not attempt to seal or cut over clips or staples as incomplete seals will be formed. Contact between an active electrode and any metal objects may result in alternate site burns or incomplete seals.
 - The surfaces of the jaws may remain hot enough to cause burns after the RF current is deactivated.
 - Inadvertent activation or movement of the activated instrument outside of the field of vision may result in injury to the patient.
 - Do not activate the LigaSure system in an open-circuit condition. Activate the system only when the instrument is in direct contact with the target tissue to lessen the possibility of unintended burns.
 - Do not activate the instrument while instrument jaws are in contact with, or in close proximity to, other instruments including metal cannulas, as localized burns to the patient or physician may occur.
3. Squeeze the lever until it latches in place.

Precaution

- Do not activate the LigaSure system until the lever has been latched. Activating the system before latching may result in improper sealing and may increase thermal spread to tissue outside of the surgical site.
 - Do not overfill the jaws of the instrument with tissue, as this may reduce device performance.
 - Use caution during surgical procedures in which patients exhibit certain types of vascular pathology (atherosclerosis, aneurysmal vessels, etc.). For best results, apply the seal to unaffected vasculature.
 - Keep the instrument jaws (1) clean. Build-up of eschar may reduce the seal and/or cutting effectiveness. Wipe jaw surfaces and edges with a wet gauze pad as needed.
4. Relax the hand on the lever once latched. The device will maintain the latched position.

Notice – Do not apply additional hand force to the lever during sealing to ensure proper function.

5. The LigaSure system can be activated by either of the following two methods:

- Press and hold the activation button (5) on the back of the instrument.

or

- By using an appropriate footswitch associated with the generator.

Refer to the generator user's guide for footswitch instructions.

Note: If both the activation button and footswitch pedal are activated during the same seal cycle, the vessel sealing system delivers energy from the activation source it detects first.

A continuous tone sounds to indicate the activation of RF energy. When the activation cycle is complete, a two-pulsed Seal-Cycle-Complete tone sounds and RF output ceases.

A tone with multiple pulses indicates that the seal cycle was not completed. Refer to the Troubleshooting section on page 6 for possible causes and corrective actions. Do not cut tissue until you have verified that there is adequate seal.

Notice – Keep the activation button dry and clean.

6. Release the activation button on the instrument or the footswitch pedal when the seal cycle is complete.
7. To seal adjacent tissue, overlap the edge of the existing seal. The second seal should be distal to the first seal to increase seal margin.

Cutting Tissue

Warning

- Prior to cutting the seal, inspect the vessel or tissue to ensure proper sealing.
- Energy-based devices, such as ES pencils or ultrasonic scalpels that are associated with thermal spread,

should not be used to transect seals.

Notice – Do not engage the cutting mechanism over clips, staples, or other metal objects as damage to the cutter may occur.

1. To engage the cutting mechanism, pull the cutting trigger (9) completely back toward the body of the instrument.

Important: If the cutting trigger does not automatically return to position, unlatch and open the lever to manually return the cutting trigger.

2. Open the jaws by squeezing the lever until it unlocks, then push it completely forward.

Warning

If the lever cannot be unlatched following use, open the device by forcing the lever forward from the handle. The device will no longer function properly and must be discarded.

Cleaning the Instrument During Use

Warning

- Inspect the instrument jaws prior to cleaning to ensure the blade is not deployed.
- Do not activate the instrument or cutting trigger while cleaning the jaws. Injury to operating room personnel may result.

1. Wipe jaw surfaces and edges with a wet gauze pad as needed.

Notice

- Do not attempt to clean the instrument jaws by activation the instrument on wet gauze. Product damage may occur.
- Do not clean the instrument jaws with a scratch pad or other abrasives.

Troubleshooting

Alert Situations

When an alert condition occurs, energy delivery is terminated, but will be available immediately after the alert condition has been corrected.

Troubleshooting Information	
The following is a list of troubleshooting suggestions for situations encountered when using the instrument with compatible Covidien vessel sealing generators. For details on specific situations, refer to the corresponding generators user’s guide or the generator quick reference guide.	
Alert Situations	When an alert condition occurs, energy delivery stops, the generator produces a sequence of pulsed tones, and an alert will be displayed on the generator. Do Not Cut the Vessel. The user should inspect the seal site and instrument before proceeding. After the alert condition has been corrected, energy delivery will be immediately available.
Troubleshooting Steps	<ol style="list-style-type: none"> 1. Release the footswitch pedal or activation button, if still engaged. 2. Open the instrument jaws and inspect for a successful seal. 3. Follow the suggested corrective actions on the generator screen, the generator quick reference card, or in the generator user’s guide. 4. If possible, reposition the instrument and regrasp tissue in a location that overlaps the previous seal, then reactivate the seal cycle.
Reasons for Alert	Too little tissue between the jaws – The user is grasping thin tissue or not enough tissue; open the jaws and confirm that a sufficient amount of tissue is inside the jaws. If necessary, increase the thickness of tissue that is grasped and reactivate the seal cycle.

	<p>Too much tissue between the jaws – The user is grasping too much tissue; open the jaws, reduce the amount of tissue that is grasped, and reactivate the seal cycle.</p> <p>Activating on metal object – Avoid grasping objects, such as staples, clips, or encapsulating sutures in the jaws of the instrument.</p> <p>Dirty jaws – Use a wet gauze pad to clean surfaces and edges of instrument jaws.</p> <p>Excess Fluids in the Surgical Field – Minimize or remove excess fluids from around the instrument jaws.</p> <p>Activation switch released before seal complete tone – The footswitch or activation button was released before the seal cycle was complete.</p> <p>Maximum seal cycle time has been reached – The system needs more time and energy to complete the seal cycle.</p>
--	--

After Surgery

- Place the instrument in the appropriate collection container after use according to hospital policy. **Do not resterilize.**

Pre-Clinical Study Notice

There is no animal data qualified to predict the effectiveness of this device in sealing vessels containing atherosclerotic plaque.

Clearance of the device in the United States was not based on human clinical testing; instead, performance of the device was established in a chronic in vivo porcine model. The results showed that no animals studied experienced any hemostatic complications related to the device during the 21-day survival period. The evaluation was performed on a variety of tissue types and vessels, up to and including 7 mm, to demonstrate effective sealing.

Vessel Type	Vessel/Tissue Name	Size Range (mm)
A/V Bundle	Short Gastric	1-5
	Splenic	5
	Ovarian Pedicle	3-7
	Small Bowel Mesentery	≤ 1
Isolated Vein	Gastrosplenic	1-3
	Splenic	3-7
	Renal	5-7
	Phrenicoabdominal	1-6
	Ovarian Pedicle	3
	Jugular	6-7
Isolated Artery	Gastrosplenic	2-3
	Splenic	4-6
	Renal	2-7
	Phrenicoabdominal	1-3
	Carotid	5-7

Transport and Storage Conditions

- An ambient temperature range of -18°C to 60°C.
- Do not expose to relative humidity below 15% or above 90%.

Operating Conditions

An ambient temperature range of +10°C to +40°C.

Warranty

Reprocessed products

Stryker warrants all reprocessed products, subject to the exceptions provided herein, to be free from defects in reprocessing and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for one use in accordance with the instructions for use of such product.

Products for which Stryker is the original manufacturer

Stryker warrants all products for which it is the original manufacturer, subject to the exceptions provided herein, to be free from defects in design, materials and workmanship and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for a period of one year from the date of purchase.

General warranty terms applicable to all products

To the fullest extent permitted by law, the express warranty set forth herein is the only warranty applicable to the products and is expressly in lieu of any other warranty by Stryker, expressed or implied, including, but not limited to, any implied warranty or merchantability or fitness for a particular purpose. In no event will Stryker's liability arising in connection with the sale of the product (whether under the theories of breach of contract, tort, misrepresentation, fraud, warranty, negligence, strict liability or any other theory of law) exceed the purchase price, current market value or residual value of the products, whichever is less. Stryker shall not be liable for indirect, special, incidental, punitive, or consequential damages resulting from any breach of warranty or under any other legal theory.

This warranty shall apply only to the original end-user purchaser of products directly from Stryker or a Stryker authorized distributor. This warranty may not be transferred or assigned without the express written consent of Stryker.

This warranty does not apply to: (1) products that have been misused, neglected, modified, altered, adjusted, tampered with, improperly installed or refurbished; (2) products that have been repaired by any person other than Stryker personnel without the prior written consent of Stryker; (3) products that have been subjected to unusual stress or have not been maintained in accordance with the instructions in the user manual or as demonstrated by a Stryker representative; (4) products on which any original serial numbers or other identification marks have been removed or destroyed; or (5) products that have been repaired with any unauthorized or non-Stryker components.

If a valid warranty claim is received within thirty (30) days of the expiration of the applicable warranty period, Stryker will, in its sole discretion: (1) replace the product at no charge with a product that is at least functionally equivalent to the original product or (2) refund the purchase price of the product. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property. In any event, Stryker's liability for breach of warranty shall be limited to the replacement value of the defective or non-conforming part or component.

If Stryker determines in its reasonable discretion that the claimed defect or non-conformance in the product is excluded from warranty coverage as described hereunder, it will notify the customer of such determination and will provide an estimate of the cost of repair of the product. In such an event, any repair would be performed at Stryker's standard rates.

Products and product components repaired or replaced under this warranty continue to be warranted as described herein during the initial applicable warranty period or, if the initial warranty period has expired by the time the product is repaired or replaced, for thirty (30) days after delivery of the repaired or replaced product. When a product or component is replaced, the item provided in replacement will be the customer's property and the replaced item will be Stryker's property. If a refund is provided by Stryker, the product for which the refund is provided

Reprocessed LigaSure Blunt Tip Sealer/Divider Without Nano-coating (Model LF1837)

Page 10 of 10

must be returned to Stryker and will become Stryker's property.

The OEM information listed on the label is provided as device ID prior to reprocessing and may contain the trademarks of unrelated third parties that do not sponsor this device.

Sterilization: This product and its packaging have been sterilized with ethylene oxide gas (EtO). Even though the product then is processed in compliance with all applicable laws and regulations relating to EtO exposure, Proposition 65, a State of California voter initiative, requires the following notice:

Warning: This product and its packaging have been sterilized with ethylene oxide. The packaging may expose you to ethylene oxide, a chemical known to the State of California to cause cancer or birth defects or other reproductive harm.

Covidien™ is a registered trademark of Covidien AG.
ForceTriad™ and LigaSure™ and are trademarks of a Covidien company.

VSD EL10107 Rev. E 05/2024