



AXS

Axial stability screw

1.2mm and 1.7mm

AXS


Greater ease in handling in comparison to predicate devices¹ by implementation of the following features²:

- Improved screw pick-up performance
- Improved off-axis stability during insertion (fig. 1)
- Lower profile height of screw head
- Improved insertion of self-drilling screws



(fig. 1)

Vertical



1.2mm Self-drilling screws

Product #	Description
56-12903	1.2x3mm
56-12904	1.2x4mm

1.7mm Self-drilling screws

Product #	Description
56-17903	1.7x3mm
56-17904	1.7x4mm
56-17905	1.7x5mm
56-17906	1.7x6mm

1.2mm Self-tapping screws

Product #	Description
56-12003	1.2x3mm
56-12004	1.2x4mm
56-12005	1.2x5mm
56-12006	1.2x6mm
56-12007	1.2x7mm
56-12008	1.2x8mm
56-12010	1.2x10mm
56-12012	1.2x12mm

1.7mm Self-tapping screws

Product #	Description
56-17003	1.7x3mm
56-17004	1.7x4mm
56-17005	1.7x5mm
56-17006	1.7x6mm
56-17007	1.7x7mm
56-17008	1.7x8mm
56-17010	1.7x10mm
56-17012	1.7x12mm

1.4mm Emergency screws

Product #	Description
56-14003	1.4x3mm
56-14005	1.4x5mm
56-14007	1.4x7mm
56-14009	1.4x9mm

1.7mm Locking screws

Product #	Description
56-17504	1.7x4mm
56-17505	1.7x5mm
56-17506	1.7x6mm

1.2/1.7mm Screwdriver blades

Product #	Description
62-12172	42mm - short
62-12174	82mm - standard
62-12177	112mm - long

1.9mm Emergency screws

Product #	Description
56-19003	1.9x3mm
56-19005	1.9x5mm
56-19007	1.9x7mm
56-19009	1.9x9mm

¹ Tested in comparison to the standard screws and screwdriver blades from the Universal 2 System

² Data on file

Head and Neck

This document is intended solely for the use of healthcare professionals.

A healthcare professional must always rely on his or her own professional clinical judgment when deciding whether to use a particular product when treating a particular patient. Stryker does not dispense medical advice and recommends that healthcare professionals be trained in the use of any particular product before using it in surgery.

The information presented is intended to demonstrate the breadth of Stryker product offerings. A healthcare professional must always refer to the package insert, product label and/or instructions for use before using any Stryker product.

Products may not be available in all markets because product availability is subject to the regulatory and/or medical practices in individual markets. Please contact your Stryker representative if you have questions about the availability of Stryker products in your area.

Stryker Corporation or its divisions or other corporate affiliated entities own, use or have applied for the following trademarks or service marks: Stryker. All other trademarks are trademarks of their respective owners or holders.

The products depicted are CE marked in accordance with applicable EU Regulations and Directives.



CMFFL8016543EN
DLS 01/2018
2018-16625

Copyright © 2018 Stryker
stryker.com



Stryker Leibinger
Bötzinger Straße 41
79111 Freiburg/Germany



0197