

# Sonopet iQ

## Ultrasonic Aspirator System

### Tip Guide



Bone Cutting

#### 12cm iQ Micro Claw

5500-25B-315  
4/box

L 4.8in (121mm) CL 1.6mm CW 2mm



#### 12cm iQ Claw

5500-25B-318  
4/box

L 4.8in (121mm) CL 1.7mm CW 2.8mm



#### 12cm iQ 120 Claw

5500-25B-317  
4/box

L 4.8in (121mm) CL 2.58mm CW 3.15mm



#### 12cm iQ Apex 360

5500-25B-312  
4/box

L 4.8in (121mm) CL 2.58mm CW 3.15mm



#### 20cm iQ Apex 360

5500-25B-302  
4/box

L 7.9in (200mm) CL 2.5mm CW 3.15mm



#### 18cm iQ Micro Claw

5500-25B-303  
4/box

L 7.2in (182.8mm) CL 1.6mm CW 2mm



#### 18cm iQ Apex Micro Claw

5500-25B-319  
4/box

L 7.2in (182.8mm) CL 1.6mm CW 2mm



#### 11cm iQ Knife

5500-25B-107  
4/box

L 4.2in (106mm) CL 12.5mm CW 0.83mm



#### 11cm iQ Apex Knife

5500-25B-114  
4/box

L 4.2in (106mm) CL 12.5mm CW 0.83mm



L = Length CL = Cutting length CW = Cutting width

Soft Tissue

#### 12cm iQ Micro

5500-25S-309  
4/box

L 4.7in (120mm) OD 1.6mm ID 1.20mm



#### 12cm iQ Standard

5500-25S-307  
4/box

L 4.5in (115mm) OD 1.92mm ID 1.5mm



#### 12cm iQ Large

5500-25S-308  
4/box

L 4.7in (112.1mm) OD 2.49mm ID 2.0mm



#### 20cm iQ Standard

5500-25S-301  
4/box

L 8.1in (205mm) OD 1.92mm ID 1.5mm



#### 12cm iQ Barracuda

5500-25S-310  
4/box

L 4.5in (115mm) OD 1.92mm ID 1.5mm



#### 12cm iQ Serrated Large

5500-25S-316  
4/box

L 4.6in (115.7mm) OD 2.51mm ID 2.0mm



L = Length OD = Outer diameter ID = Inside diameter

All tips ship with associated irrigation sleeves.



### Neurosurgical

This document is intended solely for the use of healthcare professionals.

A healthcare professional must always rely on his or her own professional clinical judgment when deciding whether to use a particular product when treating a particular patient. Stryker does not dispense medical advice and recommends that healthcare professionals be trained in the use of any particular product before using it in surgery.

The information presented is intended to demonstrate the breadth of Stryker product offerings. A healthcare professional must always refer to the package insert, product label and/or instructions for use before using any Stryker product.

Products may not be available in all markets because product availability is subject to the regulatory and/or medical practices in individual markets. Please contact your Stryker representative if you have questions about the availability of Stryker products in your area.

Stryker Corporation or its divisions or other corporate affiliated entities own, use or have applied for the following trademarks or service marks: Stryker, Barracuda and Sonopet iQ. All other trademarks are trademarks of their respective owners or holders.

The products depicted are CE marked in accordance with applicable EU Regulations and Directives.

SDL 04/2020  
2020-24224

Copyright © 2020 Stryker  
stryker.com



Stryker  
4100 East Milham Avenue  
Kalamazoo, MI 49001 USA  
Tel: (+1) 269 323 7700  
Fax: (+1) 800 999 3811