Reprocessed by



**Sustainability Solutions** 

## Instructions for Use Reprocessed Pressure Infusor Bag

## **Reprocessed Device for Single Use**

Caution: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

- NON-STERILE
- Exposed to Ethylene Oxide (EO) gas
- Not made with natural rubber latex

### **Explanation of Symbols**



Date of Reprocessing

Do Not Reuse



See Instructions For Use

Stryker Sustainability Solutions, Inc. © 2019 1810 W Drake Dr. Tempe AZ, 85283 sustainability.stryker.com 888.888.3433

#### **Reprocessed Pressure Infusor Bag Description**

The Reprocessed Pressure Infusor Bag consists of a bag with a bladder, an envelope for holding the fluid bag, a pressure gauge, and tubing with an inflation bulb and a stopcock. Once the fluid bag is loaded in the envelope, the bag will then be inflated to the desired pressure.

#### Indications for Use

Reprocessed Pressure Infusor Bags are intended to be used to apply pressure to a sealed bag of sterile fluid to assist in the infusion of the fluid.

#### **Contraindications for Use**

None known.

#### Warnings

- Prior to infusion, it is recommended that the administration set be properly primed and residual air removed from the fluid bag and tubing to minimize the risk of air embolism.
- Make sure pressure gauge can move freely and is clear of all obstructions.
- Check the infusor pressure periodically during use.

#### Precautions

- Read instructions prior to use
- Ensure adequate inflation pressure during use.
- Do not use if package is damaged.

#### **Directions for Use**

- 1. Insert intravenous bag between netting and the air bladder.
- 2. Hang intravenous bag on the hook located on the back of the air bladder.
- 3. Use IV pole loop to hang pressure infusor bag from the IV pole.
- 4. To inflate, turn stopcock handle to point toward the open stopcock vent.
- 5. Pump inflation bulb until the pressure gauge indicates the desired pressure.
- 6. IMPORTANT: To maintain pressure in the bag and ensure a leak proof seal, turn stopcock handle straight up towards the air bladder to the maintain pressure position.
  - a. Note: Precise positioning of the stopcock is important or bag could lose pressure.
- 7. To deflate, turn stopcock handle straight down towards the inflation bulb.

#### **Storage and Handling**

- Storage conditions: Room temperature, ventilation, dry, away from light.
- Transportation conditions: Do not expose to excessive heat or humidity during transportation.

#### Warranty

#### Reprocessed Products

Stryker warrants all reprocessed products, subject to the exceptions provided herein, to be free from defects in reprocessing and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for one use in accordance with the instructions for use of such product.

# STRYKER SHALL NOT BE LAIBLE FOR ANY DAMAGES TO THE EXTENT CAUSED BY ANY DEFECT IN MATERIAL, WORKMANSHIP OR DESIGN BY THE ORIGINAL MANUFACTURER OF THE PRODUCT OR ANY ACT OR OMISSION OF THE ORIGINAL MANUFACTURER OF THE PRODUCT.

#### Products for which Stryker is the Original Manufacturer

Stryker warrants all products for which it is the original manufacturer, subject to the exceptions provided herein, to be free from defects in design, materials and workmanship and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for a period of one year from the date of purchase.

#### **General Warranty Terms Applicable to All Products**

TO THE FULLEST EXTENT PERMITTED BY LAW, THE EXPRESS WARRANTY SET FORTH HEREIN IS THE ONLY WARRANTY APPLICABLE TO THE PRODUCTS AND IS EXPRESSLY IN LIEU OF ANY OTHER WARRANTY BY STRYKER, EXPRESSED OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OR MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. IN NO EVENT WILL STRYKER'S LIABILITY ARISING IN CONNECTION WITH THE SALE OF THE PRODUCT (WHETHER UNDER THE THEORIES OF BREACH OF CONTRACT, TORT, MISREPRESENTATION, FRAUD, WARRANTY, NEGLIGENCE, STRICT LIABILITY OR ANY OTHER THEORY OF LAW) EXCEED THE PURCHASE PRICE, CURRENT MARKET VALUE OR RESIDUAL VALUE OF THE PRODUCTS, WHICHEVER IS LESS. STRYKER SHALL NOT BE LIABLE FOR INDIRECT, SPECIAL, INCIDENTAL, PUNITIVE OR CONSEQUENTIAL DAMAGES RESULTING FROM ANY BREACH OF WARRANTY OR UNDER ANY OTHER LEGAL THEORY.

This warranty shall apply only to the original end-user purchaser of products directly from Stryker or a Stryker authorized distributor. This warranty may not be transferred or assigned without the express written consent of Stryker.

This warranty does not apply to: (1) products that have been misused, neglected, modified, altered, adjusted, tampered with, improperly installed or refurbished; (2) products that have been repaired by any person other than Stryker personnel without the prior written consent of Stryker; (3) products that have been subjected to unusual stress or have not been maintained in accordance with the instructions in the user manual or as demonstrated by a Stryker representative; (4) products on which any original serial numbers or other identification marks have been removed or destroyed; or (5) products that have been repaired with any unauthorized or non-Stryker components.

If a valid warranty claim is received within thirty (30) days of the expiration of the applicable warranty period, Stryker will, in its sole discretion: (1) replace the product at no charge with a product that is at least functionally equivalent to the original product or (2) refund the purchase price of the product. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property. In any event, Stryker's liability for breach of warranty shall be limited to the replacement value of the defective or non-conforming part or component.

If Stryker determines in its reasonable discretion that the claimed defect or non-conformance in the product is excluded from warranty coverage as described hereunder, it will notify the customer of such determination and will provide an estimate of the cost of repair of the product. In such an event, any repair would be performed at Stryker's standard rates.

Products and product components repaired or replaced under this warranty continue to be warranted as described herein during the initial applicable warranty period or, if the initial warranty period has expired by the time the product is repaired or replaced, for thirty (30) days after delivery of the repaired or replaced product. When a product or component is replaced, the item provided in replacement will be the customer's property and the replaced item will be Stryker's property. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property.

The OEM information listed on the label is provided as device ID prior to reprocessing and may contain the trademarks of unrelated third parties that do not sponsor this device.

This product and its packaging have been exposed to ethylene oxide gas (EtO). Even though the product then is processed in compliance with all applicable laws and regulations relating to EtO exposure, Proposition 65, a State of California voter initiative, requires the following notice:

Warning: This product and its packaging have been exposed to ethylene oxide. The packaging may expose you to ethylene oxide, a chemical known to the State of California to cause cancer or birth defects or other reproductive harm.

PIB EL10101 Rev. B 09/2019